

TELEPHONE: 0414-696333
GENERAL LINE: 0485-660639
FAX: 041425342
E-MAIL: secretary@atomiccouncil.go.ug
Website: www.atomiccouncil.go.ug



ATOMIC ENERGY COUNCIL
PLOT 29/33 AMBER HOUSE,
KAMPALA ROAD,
P. O. BOX 7044,
KAMPALA.

CONDITIONS FOR DECOMMISSIONING A RADIOACTIVE SOURCE

These conditions have been issued in line with the Atomic Energy Regulations, 2012.

1. The facility shall comply with the requirements of the Atomic Energy Act, 2008 and the Atomic Energy Regulations, 2012, related to radiation protection and safety.
2. The facility shall follow the activities as indicated in the decommissioning plan submitted to Council.
3. The facility shall ensure that the temporary storage area where the source is to be kept meets the storage safety and security requirements provided by the Act and Regulation.
4. The facility shall implement the Security Plan during decommissioning and storage of the Radioactive Source.
5. The facility shall carry out regular area surveys around the storage area of the source
6. The facility shall ensure the safety of all radiation persons involved in the decommissioning from any possible radiological hazards.
7. The facility shall report to Council immediately any incidents or accidents that may arise during and after the decommissioning process, following the emergency plan that was included in the decommissioning plan.
8. The facility shall ensure that arrangements are in place to return the sealed source to the manufacturer for final disposal as indicated in the decommissioning plan.
9. The facility shall make a final safety and security assessment of the storage area at the end of the decommissioning process.
10. The facility shall submit a final decommissioning report to Council within 30 days from the date of completion of the decommissioning process.

Signed by:

.....
Deogratias Noah Luwalira
SECRETARY/CHIEF EXECUTIVE OFFICER

.....
Date Stamp