



General Terms and Conditions of Dosimetry Service

Section 13 (1) of the Atomic Energy Act, Cap. 154 mandates the Council to operate and maintain a dosimetry service.

In order to provide efficient and effective dosimetry services, the Atomic Energy Council (AEC) has established terms and conditions.

1. AEC thermoluminescent dosimeters (TLDs) shall remain the property of the Council at all times.
2. Facilities shall apply for dosimetry services in the prescribed form (Attach a template of the application form).
3. The TLDs shall be worn by personnel at facilities for three months from the date of issue or as required by AEC.
4. All issued TLDs whether used or unused shall always be returned to AEC for reading, calibration and annealing after the wearing period stipulated in paragraph 3 above. The returned TLDs shall be accompanied with the list of names of the wearers.
5. In case a facility returns less TLDs than those which were issued to it, it will not be issued new TLDs not until it returns all the missing TLDs or pays for the missing TLDs.
6. The damaged or lost TLDs shall be replaced by either paying an amount worth Ugx. 700,000/= for each TLD (Uganda Shillings Seven Hundred Thousand Only) as per the Atomic Energy (Dosimetry Services Fees Notice, 2024) or by returning the compatible TLDs. In case of the latter, Council shall determine the specifications.
7. AEC shall explore remedies under the Atomic Energy Act, Cap. 154 and regulations made thereunder where facilities fail to return the TLDs within the stipulated time.
8. Facilities shall notify AEC of lost or damaged TLDs immediately it comes to their attention.

9. Facilities shall pay Ugx. 10,000 (Uganda Shillings Ten Thousand Only) per TLD for reading, annealing, and calibration of the TLDs upon return of the TLDs as per Schedule 6 of the Atomic Energy Regulations, 2012.
10. Facilities shall not be issued with the dosimetry reports and new TLDs without making payment for the reading, annealing, and calibration of the returned TLDs.
11. Each TLD shall be assigned to and used by one (01) radiation worker per facility during the wearing period.
12. TLD(s) shall not be used by different facilities during the same wearing period.
13. It is the responsibility of the facilities to pick their dosimetry results from Council.